


ELECTRON AND ION OPTICS GUNS / SOURCES

5 eV to 100 keV

High Brightness Electron Sources
High Brightness Ion Sources
UHV Electrostatic Optics, Power Supplies
Charge Neutralization Systems
Accessories for SEM, MBE, ESD, SIMS, AUGER, e^- Lithography
Custom UHV Apparatus



UP TO 20 keV AVAILABLE ON 70 mm (2 $\frac{3}{4}$ in.) CONFLAT

KIMBALL PHYSICS INC.



KIMBALL HILL ROAD
WILTON, NH 03086

(603) 878-1616



UHV ELECTRON & ION SOURCES/OPTICS/SYSTEMS

Circle number 77 on Reader Service Card

ANNOUNCING THE IBM-PCA



**Personal
Computer
Analyzer**

The IBM-PCA provides the ultimate solution to your multichannel pulse height analyzer requirements. By building the hardware on a card which fits inside the computer, we've eliminated the need for an external buffer and expensive NIM system.

No more lengthy data transfers. The information is already resident in the computer as part of the IBM memory map.

Both hardware and software are compatible with the IBM-PC, XT and AT systems, the PCA occupies only a single slot yet contains the ADC, single channel analyzer, 24-bit dynamic memory and a 50MHz multichannel scaler.

REPRESENTATIVES WORLDWIDE

761 EMORY VALLEY RD., OAK RIDGE, TN 37830-2561
TELEPHONE 615-482-4041, TLX 557-482



Circle number 78 on Reader Service Card

College and his PhD in 1972 from the University of Pittsburgh, where he remained for a year as a postdoctoral research associate. He joined Brooklyn College in 1973 as an instructor, became an assistant professor in 1974 and an associate professor in 1978.

Thomas made major contributions to the application of the Glauber scattering theory to atomic physics, including reductions to closed form of the Glauber amplitudes both for collisions of charged particles with hydrogen atoms and for excitation of hydrogen-like ions by structureless charged particles. In addition he wrote a lucid review article on applications of the Glauber approximation to atomic collisions. Thomas had a remarkable knack for obtaining analytical results for even the most unwieldy-looking integrals. He used this talent to help develop the Glauber theory into a useful tool for atomic-scattering calculations. These skills also made him an invaluable resource to his colleagues.

Thomas was, furthermore, a very gifted teacher. His creativity and enthusiasm were apparent whether he was teaching the most elementary courses or advanced graduate courses, and he was popular with his students.

As a long-time member of the admissions committee of City University's multicampus doctoral program in physics, and as its chairman for two years, he played an important role in the program's ongoing success. His spirit was always buoyant and optimistic and his sense of humor was keen. His colleagues and friends will remember him as one who gave selflessly of his time and energy to others and one who was refreshingly independent-minded and free-thinking.

VICTOR FRANCO
ALVIN HALPERN
Brooklyn College

Henry Horstman

Henry McKnight Horstman, professor of astronomy at the University of Bologna, died on 9 November 1984. A native of Pennsylvania, he obtained his PhD in physics from Carnegie Tech in 1962 while working under Sergio De Benedetti. He spent seven years at the NASA Goddard Space Flight Center, experimenting with balloon- and rocket-borne detectors designed to study the newly discovered x-ray emission from celestial objects. He came to the University of Bologna in 1968 with his wife, Elena, who became a research scientist with the National Research Council in Italy. They collaborated on several research projects, including a "gamma camera" designed to image hard x rays in the 30-300-keV range and intended for astronomical observations. The

concept of the camera was derived from medical x-ray equipment. The camera was test-flown in 1983 and scheduled for flight in the summer of 1985.

Horstman was a complete scientist, capable not only of building instruments, but also of interpreting his and other's experimental results and doing theoretical work. His research ideas were unconventional and always provocative; his spirit was always high. His strong sense of humor and rigorous honesty gave him a special place in the laboratory. A devoted teacher and a generous and patient collaborator, he will be missed by those who knew him—friends and students alike.

G. G. C. PALUMBO

G. PIZZICHINI


*Consiglio Nazionale Delle Ricerche
Istituto Tecnologie e Studio delle
Radiazioni Extraterrestri
Bologna, Italy*

Gopinath Kartha

Gopinath Kartha died on 17 June 1984. He was born in Shertallay, Kerala State, India, in 1927, and received his PhD in physics from the University of Madras in 1953. In 1959 he joined David Harker in the Protein Structure Project at the Polytechnic Institute of Brooklyn, and later that year moved to Roswell Park Memorial Institute when Harker became director of the biophysics department there. Kartha held appointments as research professor in the Roswell division of SUNY at Buffalo and as research professor at Niagara University.

Kartha's primary research interest was the development and application of x-ray crystallographic techniques for investigating biological molecules. He developed experimental and theoretical techniques for solving the phase problem in the study of macromolecules, using large computers. In addition, he was active in training and guiding students in x-ray crystallographic research.

KARTHA



Kartha's scientific career was marked by a series of fine achievements; the earliest of these was the determination of the structure of collagen. The triple-helical structure of collagen, which Kartha and Gopalasamudram Ramachandran devised, can truly be said to form the foundation for all subsequent understanding of this most important protein. The combination of the x-ray derived structure and the chemically derived amino-acid sequence is a classic example of a fruitful combination of physics and chemistry in molecular biology.

Kartha then published important theoretical work on the determination of protein structure by isomorphous replacement and anomalous scattering. He was instrumental in resolving the structure of the protein ribonuclease, a pioneering work in this field. During this period he also published what is known as Kartha's method for the determination of scale factors.

In more recent years Kartha again turned his attention to molecules of medical importance. He made pioneering studies of steroid hormones; he elucidated the structures of antibiotics, antitumor agents, cardiac glycosides, amino acids, nucleic-acid components and vitamin derivatives. Most recently he worked on the structures of cyclic peptides and their complexes with metal ions, especially calcium.

Kartha's career was marked by a generous helpfulness to students and colleagues. He was devoted to the joy of research and scholarship.

JAKE BELLO

Roswell Park Memorial Institute

Sigmund W. Leifson

Sigmund W. Leifson of the University of Nevada in Reno died on 10 February 1984; he was 86. Leifson was born in North Dakota, but moved with his family to Norway in 1900, then returned in 1914. As a graduate student at the University of California at Berkeley he was one of the earliest researchers in spectroscopy of the vacuum ultraviolet. L. W. Hartman persuaded him to join the staff of the University of Nevada in 1925. Leifson, a quiet and unpretentious man, was active in improving the scope and quality of physics instruction there, and after World War II began adding staff. He gave much time and thought to careful planning. When he came to the University—and until after the Korean War—the physics and chemistry departments shared a single two-story building. When Leifson retired in 1963, each occupied a large, modern building and they shared a lecture and preparation complex. □



PEARSON Wide Band, Precision CURRENT MONITOR

With a Pearson current monitor and an oscilloscope, you can measure pulse or ac currents from milliamperes to kiloamperes, in any conductor or beam of charged particles, including those at very high voltage levels.

This monitor is physically isolated from the circuit. It is a terminated current transformer whose output voltage precisely follows the current amplitude and waveshape. A typical model gives an amplitude accuracy of +1%, -0%, 20 nanosecond rise time, droop of 0.5% per millisecond, and a 3 db bandwidth of 1 Hz to 35 MHz. Other models feature 2 nanosecond rise time, or a droop as low as 1% per second.

Whether you wish to measure current in a conductor, an electron device, or a particle accelerator, it is likely that one of our off-the-shelf models (ranging from 1/2" to 10 3/4" ID) will do the job. We also provide custom designs to meet individual specifications.

Contact us and we will send you engineering data.

PEARSON ELECTRONICS, INC.

1860 Embarcadero Road
Palo Alto, Calif. 94303, U.S.A.
Telephone (415) 494-6444
Telex 171-412